



MINUTES OF MEETING

For

Pre-Bid Meeting for the Purchase of Biomedical Equipment Of Jinnah Burn & Reconstructive Surgery Centre, Lahore

A Pre-Bid Meeting was held on 10.02.2026 at 10:30 AM in the office of F-08 Private Clinic, Jinnah Burn & Reconstructive Surgery Centre (JB&RSC), Lahore, to discuss and evaluate clarifications regarding the **Purchase of Biomedical Equipment** through E-PADS, in reference to Meeting Notice No. ___ dated 09.02.2026.

The meeting was chaired by the Convener TAC, Jinnah Burn & Reconstructive Surgery Centre, Lahore, in the presence of the following committee members:

1.	Prof. Uzma Hamza HOD Ophthalmology, Unit-II JHL	Chairman
2.	Dr. Ata-ul-Haq Associate Prof. of Plastic Surgery In-charge Procurement, JB&RSC, Lahore	Member
3.	Dr. Lala Rukh Bangash Associate Professor of Anesthesia, JB&RSC, Lahore	Member
4.	Dr. Muhammad Amin Assistant Professor of Plastic Surgery, JB&RSC, Lahore	Member
5.	Syed Muhammad Shakir Biomedical Engineer, JB&RSC, Lahore	Member

PROCEEDINGS

The meeting commenced with the name of Almighty Allah, the Most Gracious, the Most Merciful. The Purchase Assistant briefed the committee that the tender for procurement of biomedical equipment had been duly advertised on the PPRA website as well as on the E-PADS portal. Subsequently, several firms submitted clarifications, observations, and reservations regarding various items included in the tender. Representatives of the following firms participated in the pre-bid meeting:

1. M/S Shirazi trading
2. M/S Vertex Medical
3. M/S Mediland Pakistan
4. M/S G Med (Pvt.) Ltd.
5. M/S Kasban Interantional
6. M/S Friends Traders
7. M/S Total Technologies
8. M/S Radiant Medical
9. M/S Fazal Din & Sons

Ymmi
Q. Chisht

Q.



Office of the
EXECUTIVE DIRECTOR
Jinnah Burn & Reconstructive Surgery Centre, AIMC

LAHORE

Tel: 042-99232701-2 & 042-99232730-46




10. M/S Eastern Medical
11. M/S Summit International
12. M/S Medical Equipment

The committee heard the clarifications, observations, and reservations presented by the participating firms. After detailed discussion, the committee recorded the proceedings and decisions accordingly. Minutes of the meeting are recorded as follows:



Biomedical Engineer


MEMBER
JB&RSC, LAHORE



Dr. Lala Rukh Bangash


MEMBER

ASSOC. PROFESSOR. OF ANESTHESIA,
JB&RSC, LAHORE




Dr. Muhammad Amin

MEMBER
ASSISTANT. PROFESSOR. JB&RSC, LAHORE



Dr. Ata-ul-Haq

MEMBER
INCHARGE PROCUREMENT/ASSOCIATE PROF.
JB&RSC, LAHORE



Prof. Uzma Hamza
CHAIRMAN
HOD OPHTHALMOLOGY UNIT-II, JHL

#	Firm(s) Name	Item Name	Observations / Reservations	Committee Action / Decision
1.	M/s Kasban International	Hand Held Vascular Doppler	<ul style="list-style-type: none"> The firm requested to include option of separate power ON/OFF switch and Volume Control for ease of use. 	<ul style="list-style-type: none"> The committee Reviewed the observation regarding the provision of either combined or separate Power ON/OFF and volume control buttons, and recommended appropriate configuration for improved ease of use.
2.	M/s Radiant Medical		<ul style="list-style-type: none"> Requested to include Monochrome Display instead of Color display 	<ul style="list-style-type: none"> The term color has been removed from the specifications, and committee agreed to include monochrome display.
3.	M/s Eastern Medical	ICU Ventilator	<ul style="list-style-type: none"> Requested to increase the battery backup for 04 hours instead of 02 hours. 	<ul style="list-style-type: none"> The committee rejected the proposal to increase battery backup duration, stating that 02 Hours battery backup is sufficient, and no extension required.
4.	M/s G Med	Pneumatic Dermatome	<ul style="list-style-type: none"> The firm pointed out the Duplication of point No 1&2 	<ul style="list-style-type: none"> The committee accepted the observation and Resolved the issue after rechecking the Approved PVMs of Pneumatic Dermatome is added
5.	M/s Shirazi Trading	ICU Monitor	<ul style="list-style-type: none"> The requested that please specify screen size 12", as a 12" screen size is sufficient for clear visibility of advertised parameters. 	<ul style="list-style-type: none"> The committee did not agree to change the screen size, as the required screen size remains 15 inches as per approved specifications.
6.	M/s G Med	Pneumatic Drill System	<ul style="list-style-type: none"> The firm requested to remove the Multipurpose drill in the optional and request to write Saw attachment 	<ul style="list-style-type: none"> The committee agreed that only Saw attachment will be mentioned under optional item

7. M/s Mediland	Instrument Washer Disinfector	<ul style="list-style-type: none"> The firm requested to exclude the "Automatic Trolley Recognition System/RF ID based/Bar code" due to which cost is increased. The firm also requested to exclude the OR lamp handle due to which cost is increased. 	<ul style="list-style-type: none"> The committee rejected the proposal since the specification is part of the approved PVMS, and therefore the point will not be removed.
8. M/s Mediland		<ul style="list-style-type: none"> All firms have reservations on optional item and requested to exclude from specifications "Central Dosing Pump" "preferably for setups consisting of 3 or more washer central dosing system for effective management and utilization of detergents and chemicals. (IO to specify)." Because Two dosing pumps already demanded in standard. 	<ul style="list-style-type: none"> The committee decided that the OR lamp handle carrier/basket shall be treated as an optional item may be provide locally if required.
9. M/s Total Technologies			
10. M/s Radiant Medical	Instrument Washer Disinfector		
11. M/s Summit		<ul style="list-style-type: none"> The firm have an observation that Spare 10 Jar of 04 Liter increase the cost so decrease the quantity. 	<ul style="list-style-type: none"> After reviewing the observation, the committee agreed to reduce the quantity from 10 jar to 05.
12. M/s Radiant Medical	Heavy Duty Suction Machine		

13. M/s G Med	Small Bone Drill (Electric)	<ul style="list-style-type: none"> The firm requested that Please revised the Speed range: 35000-40000rpm instead of 0-80000rpm The firm requested that Hand Piece not defined Since the specifications correspond to a pencil type drill, it is requested to clearly mention Pencil-Type Hand piece The firm requested that Specifications mention an electric system with multiple handpieces/attachments for each application, where all attachments are already listed separately. This line may kindly be removed to avoid confusion during physical inspection. 	<ul style="list-style-type: none"> The committee rejected the proposal, stating that a High range is required for plastic Surgery procedures, and therefore the existing specification will remain unchanged. The committee reviewed the observation and agreed to include the requirement for a high-speed drill, typically of pencil type, as required for surgical application.
14. M/s Shirazi Trading	Anesthesia Workstation (Low End)	<ul style="list-style-type: none"> The firm requested that please specify that change the Pendent mounted to <u>Floor Mounted</u> /<u>Mobile type</u> for Anesthesia Machine 	<ul style="list-style-type: none"> The committee resolved to remove ambiguity by standardizing the specification as "Electric hand piece compatible with multiple attachments. The committee agreed revise the specification to Floor Mounted / Mobile type.
15. M/s Mediland	Anesthesia Workstation (Medium End)	<ul style="list-style-type: none"> The firm requested to change Medium End to high end Anesthesia 	<ul style="list-style-type: none"> The committee rejected the proposal as the specified equipment already been approved from health department and BOM.

16.	M/s Hospicare	Pulse Oximeter	<ul style="list-style-type: none"> The firm requested to decrease the quantity of reusable sensor 	<ul style="list-style-type: none"> The committee agreed and revise the quantity of reusable sensor to 02 each instead of 05.
17.	M/s Total Technologies	High End Electrosurgical Unit	<ul style="list-style-type: none"> The firm raised objection against Electrosurgical unit trolley quantity mentioned 10 which is typographical mistake 	<ul style="list-style-type: none"> The committee reviewed and revise quantity from 10 to 01.
18.	M/s Radiant Medical			
19.	M/s Kasban		<ul style="list-style-type: none"> The firms requested to change Imported Mobile Cart to Local Mobile Cart due to which it may lead to the significant increase in cost. 	<ul style="list-style-type: none"> The committee reviewed and accepted the request that company may provide trolley locally.
20.	M/s Radiant Medical	ECG Machine (6-Channels)		
21.	M/s Vertex Medical	Watson Knife/Humbies knife	<ul style="list-style-type: none"> The firm requested to mentioned Cutting Width: Max 155.0 mm. or better 	<ul style="list-style-type: none"> The committee agreed to write 155mm or better.
22.	M/s Mediquips		<ul style="list-style-type: none"> The firm requested to remove the Strain & Shear wave Elastography which it may lead to the significant increase in cost. 	<ul style="list-style-type: none"> The committee agreed and removed Strain & Shear wave Elastography
23.	M/s Fuji Film	Digital Color Doppler (Low End).		
24.	M/s Shirazi Trading			